



Ref. Certif. No.

US-TÜVR-011142

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Process Monitoring System

Name and address of the applicant

CiDRA Corporate Services
50 Barnes Park North
Wallingford CT 06492, USA

Name and address of the manufacturer

CiDRA Corporate Services
50 Barnes Park North
Wallingford CT 06492, USA

Name and address of the factory

CiDRA Corporate Services
50 Barnes Park North
Wallingford CT 06492, USA

Ratings and principal characteristics

DC 22.8-25.2V;
90W;
Class III

Trademark (if any)

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

CYCLONEtrac OSM

Additional information (if necessary may also be reported on page 2)

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010
for national deviations see test report

As shown in the Test Report Ref. No. which forms part of this Certificate

31880003 001

This CB Test Certificate is issued by the National Certification Body



TÜVRheinland®

TUV Rheinland of North America, Inc.
12 Commerce Road
Newtown, CT 06470 - USA

Date: 10.05.2018

Signature:

Martin Glagla