	SONARtrac[®] Technical Note	
	TN0007	
	Subject: Setting and Changing Calibration Coefficients	
	Date: 29Jul08	
	Revision	02
	Page 1 of 2	


Question:

How do I change the sensor band calibration coefficients in the transmitter?

Answer:

The sensor band calibration coefficients are easily changed from the transmitter keypad. Proceed as follows


1. Press any key (except ENTER) on the transmitter keypad and the meter will go into Menu Mode and 'BASIC CONFIG' will appear on Line 1 of the display.
 2. Press the **ENTER** key and '→SENSOR SERIAL #' will appear on Line 2 of the display.
 3. Press the ↓ key until 'CALIBRATION' appears on Line 2 of the display.
 4. Press the **ENTER** key and '→CO' will appear on Line 3 and a value will appear on Line 4 of the display.
 5. Press the **ENTER** key and the + or – will be highlighted on Line 4. Use the ↓ and ↑ keys to change the value of each digit. Use the → or ← keys to scroll to each digit.
- Note: The values on Line 4 are displayed in Scientific Notation. When entering the values, enter them as shown on the sensor band calibration tag and set the final digits to 'e+00'.



6. Press the **ENTER** key when the values have been set and the display will update and show the new value.
7. Press the ↓ key and '→C1' will appear on Line 3 and the values for C1 on Line 4. Press the **ENTER** key to access and change the C1 value on Line 4.
8. Press the ↓ key and '→C2' will appear on Line 3 and the values for C2 on Line 4. Press the **ENTER** key to access and change the C2 value on Line 4.
9. Verify the new values are correct by using the ↓ key and scroll through the values.
10. Press the **EXIT** key to return the meter to Operational Mode

Please contact CiDRA Technical Support if there are any questions.

CiDRA Corporate Services	50 Barnes Park North	Wallingford, CT 06492
Tel. 203-265-0035	Fax. 203-294-4211	www.cidra.com

	SONARtrac[®] Technical Note	
	TN0007	
	Subject: Setting and Changing Calibration Coefficients	
	Date: 29Jul08	
	Revision	02
	Page 2 of 2	

Revision History

Rev	Date	Changed By	Approved By	Change Description
01	28Aug07	B. Markoja	M. Sapack	Initial Release
02	29Jul08	B. Markoja	B. Markoja	Revision #2