

Certificate

Certificate no.

CU 72503187 0001

License Holder:

CiDRA Corporate Services, LLC
50 Barnes Park North
Wallingford CT 06492
USA

Manufacturing Plant:

CiDRA Corporate Services, LLC
50 Barnes Park North
Wallingford CT 06492
USA

Report Number: US23TBBC 004

Client Reference: Mike Davis

Certification acc. to: UL 61010-1:2012 R11.24
CSA C22.2 No. 61010-1-12 R11.24

Product Information

Certified Product: Electronic Flow Monitoring System

Model Designation: DTX-a-bb-cd-ee (for Transmitter),
DSE-f-g-h (for Sensor Head Electronics Module),
(a, f=1; bb = AC, DC; c = 0, D, F, H, M, P;
d, e, g, h = 0-9, A-N, P-Z)

Technical Data: Rated Voltage: 1) AC 100-240V, 50/60Hz, (for systems
with DTX-a-AC-cd-ee)
2) DC 18-35V, (for systems with DTX-a-DC-
cd-ee)
Rated Power: 1),2) 25W
Protection Class: I

Remarks: Replaces Certificate CU72404084.

Appendix: 1, 1-4



Date of issue: 2025-08-08
(yr/mo/day)

TUV Rheinland of North America, Inc.
400 Beaver Brook Rd, Boxborough, MA 01719
Tel +1 (978) 266 9500, Fax +1 (978) 266-9992

www.tuv.com

 TÜVRheinland®